

*The Board of Pharmacy has received notice of the following product recall. The Board strongly encourages pharmacies to immediately review their quality assurance and recall policies and procedures to determine if any corrective action is required.*

The purpose of this email is to inform you that B. Braun Medical Inc. (BBMI) is issuing a product recall for specific batches of Administration Sets, Burette Sets, and Fluid Transfer Sets due to the potential for leakage from the male Luer lock site.

**Affected Product and Distribution Information:**

Article	Lot Number	GUDID	Description	Distribution Date Range	Region Distributed
	0061755752				
	,				
356048	0061761812	402249576955	FT1000S FLUID TRANSF SET	23.DEC.2020 - 31.MAR.2021	United States
	,				
	0061766279				
	0061746019				
418121	,	404696418345	SA3000 VENTED MICRO SPIKE ADAPTER	02.NOV.2020 - 21.JAN.2021	United States, Canada
	0061755023				
	0061747706				
	,				
	0061752064				
7A3817	,	505047452069	AMO ADMINISTRATION SET <sup>3</sup>	19.NOV.2020 - 19.JAN.2021	United States
	0061755195				
	,				
	0061790409				
7A3950	0061766049	N/A <sup>1</sup>	30ML 2PORT BURETTE ASSEMBLY	12.FEB.2021 - 22.MAR.2021	United States
	0061742253				
7A3970	,	N/A <sup>1</sup>	30ML BURETTE ASSY WITH STOPCOCK, BNS	28. OCT.2020 - 01. DEC.2020	Singapore
	0061751732				

				16.NOV.2021	
7M2650	0061790166	506011692063 5	THERASPHERE ADMIN SET <sup>4</sup>		Canada
				29.NOV.2021	
	0061745499				
	,				
	0061745500				
7M2804	,	404696482374 9	ALLERGAN UNIVERSAL FILL KIT <sup>5</sup>	15.DEC.2020 - 16.NOV.2021	United States
	0061745501				
	,				
	0061792770				
753070	0061753485	N/A <sup>1</sup>	TERUMO 60" TUBE SET W/HIGH CRACK CHK VLV	27.JAN.2021- 27.JAN.2021	United States
	0061744741				
M4341750S N	,	N/A <sup>12</sup>	BULK LARGEBORE MALE LL FITTING	15.OCT.2020 - 23.OCT.2020	United States
	0061749486				

1. Articles without GTIN/UDI numbers represent products that are contract manufactured by B. Braun Medical Inc. for other manufactures.
2. Not a finished medical device. Component only.
3. Note: AMO Administration Set is a product of Johnson & Johnson for whom B. Braun acts as a contract manufacturer.
4. Note: Therasphere Admin Set is a product of BIOCOMPATIBLES UK LTD for whom B. Braun acts as a contract manufacturer.
5. Note: Allergan Universal Fill Kits are produced by B. Braun but distributed by Allergan.