

The Board of Pharmacy has received notice of the following product recall:

Glenmark Pharmaceuticals Inc. is recalling the following product to the retail level:

Estradiol Vaginal Inserts, USP, 10 mcg, 18's and 8's

(NDC: 68462-711-88 and 68462-711-71)

NDC	Strength & pack size	Batch No	Expiration
68462-711-71	10 mcg, 8's	20180514	Apr-2020
68462-711-71	10 mcg, 8's	20180544	Apr-2020
68462-711-71	10 mcg, 8's	20180545	May-2020
68462-711-71	10 mcg, 8's	20180546	May-2020
68462-711-71	10 mcg, 8's	20180588	May-2020
68462-711-71	10 mcg, 8's	20190003	Jun-2020
68462-711-88	10 mcg, 18's	20180515	Apr-2020
68462-711-88	10 mcg, 18's	20180543	Apr-2020
68462-711-88	10 mcg, 18's	20180587	May-2020

Recall of these batches has been initiated for difficulty in pushing the plunger of the applicator.

The Board of Pharmacy strongly encourages pharmacies to immediately review their quality assurance and recall policies and procedures to determine if any corrective action is required.